

# Atlas

## EV ELISA Quick Guide

Revision 1 – November 11, 2024

The Atlas EV ELISA is an immunoassay to measure intact extracellular vesicles (EVs) in plasma, serum, urine, cell culture media, or cerebrospinal fluid (CSF). This kit contains enough reagents for 96 wells.

Supplied kit contents:

Item	Quantity	Storage
Human EV Capture Antibody 100X	80 $\mu$ L	4° C
Human EV Detection Antibody 100X	80 $\mu$ L	4° C
Sample Diluent	50 mL	4° C
Wash Buffer 10X	25 mL	4° C or Room temperature (RT)
TMB Solution	12 mL	4° C
Stop Solution	12 mL	4° C or RT
Pre-Coated 96-Well Microplate	96-well	4° C
Plate Sealers	3	RT
EV standard	25 $\mu$ L	-20° C

Materials needed:

- Microplate reader (450 nm absorbance)
- Plate Shaker
- Low-bind Eppendorf tubes
- Lab-grade deionized water
- 8 channel pipettor (300  $\mu$ L) or a plate washer (recommended)

About the kit:

- The kit provides enough reagents for 96 wells (12 x 8 well strips).
- Capture and detection antibodies have only been tested for stability in the 100X formulations.
- This assay is for research use only, not for diagnostic procedures.

**Reagent preparation:**

- Equilibrate reagents to room temperature (18–25°C) before use.
- Briefly spin down capture and detection antibody stocks.
- The volumes below are sufficient for one strip (8 wells); adjust volumes as needed.
- Prepare only the required volume of reagents for each experiment.

**1X Wash Buffer:**

Prepare 1X Wash Buffer by diluting Wash Buffer 10X with deionized water.

To make 25 mL of 1X Wash Buffer, combine 2.5 mL Wash Buffer 10X with 22.5 mL deionized water. Mix thoroughly and gently.


**Antibody Cocktail:**

Prepare Antibody Cocktail by diluting the capture and detector antibodies in Sample Diluent.

To make 500 µL of the Antibody Cocktail, combine 5 µL of Capture Antibody 100X and 5 µL of Detector Antibody 100X with 490 µL of Sample Diluent. Mix thoroughly and gently.

**EV Standard:**

After preparing a  $5 \times 10^9$  particles/mL standard (Calibrator A), we recommend 2-fold serial dilutions.



	Concentration (particles / mL)	Preparation
EV Standard	$5 \times 10^{10}$	
Calibrator A	$5 \times 10^9$	15 µL of EV Standard to 135 µL of Sample Diluent
Calibrator B	$2.5 \times 10^9$	75 µL of Calibrator A to 75 µL of Sample Diluent
Calibrator C	$1.25 \times 10^9$	75 µL of Calibrator B to 75 µL of Sample Diluent
Calibrator D	$6.25 \times 10^8$	75 µL of Calibrator C to 75 µL of Sample Diluent
Calibrator E	$3.125 \times 10^8$	75 µL of Calibrator D to 75 µL of Sample Diluent
Calibrator F	$1.5625 \times 10^8$	75 µL of Calibrator E to 75 µL of Sample Diluent
Calibrator G	$7.8125 \times 10^7$	75 µL of Calibrator F to 75 µL of Sample Diluent
Blank		75 µL of Sample Diluent

**Sample preparation:**

Prepare 75 µL of each sample to be tested. Dilute samples at least 2X using Sample Diluent.

Typical Sample Dynamic Range	
Assay Range (O.D)	0.03 – 1.90
Sample Type	Dilution Range
Plasma	4X – 64X
Serum	4X – 64X
Urine	2X – 8X
CSF	2X – 32X
Culture Media	2X – 32X

Note: Spin samples at 2,000 g for 10 minutes to remove platelets and debris.

#### Assay procedure:

1. Prepare all reagents, standards, and samples as directed above.
2. Remove excess microplate strips from the plate frame, seal and return them to 4° C storage.
3. Add 50  $\mu$ L of each sample or standard to separate wells in the microplate strips.
4. Add 50  $\mu$ L of the Antibody Cocktail to each well.
5. Seal the plate and incubate for 3 hours at room temperature on a plate shaker (600 rpm) or overnight at 4°C. Protect the plate from light during incubation.
6. Wash each well five times:
  - For the first wash, add 200  $\mu$ L 1X Wash Buffer to the wells containing sample.
  - Aspirate each well.
  - Dispense 300  $\mu$ L 1X Wash Buffer. Wash buffer should remain in the wells for at least 10 seconds.
  - Aspirate each well. Ensure complete removal of liquid between washes.
  - Repeat four times for a total of 5 washes.
  - After the final wash, blot the plate against clean paper towels.
7. Add 100  $\mu$ L of TMB Substrate Solution to each well. Incubate for 10 minutes in the dark on a plate shaker at 600 rpm.
8. Add 100  $\mu$ L of Stop Solution to each well. Shake for 1 minute at 300 rpm. The solution color will change from blue to yellow.
9. Measure the absorbance at 450 nm (endpoint reading). If wavelength correction is available, set to 620 nm.

#### Data analysis:

We recommend running the samples and calibrators in duplicate. To calculate the results, average the duplicate OD readings and subtract the average OD of the zero standard. Fit the standard curve using 4PL or a linear regression on a log-log plot. Interpolate the standard curve to determine the samples' concentrations. Multiply the calculated concentration by the dilution factor to obtain the final concentration. A new standard curve should be generated with every experiment.

#### Warnings and precautions:

The Sample Diluent contains ProClin 300®, which may cause an allergic skin reaction. Avoid ingestion and contact with eyes, skin, and mucous membranes.

#### Example of a standard curve:

